

Report: SSL.19F006.IB2-L3

Issued: 19 July 2019

Page: 1 of 8

Test Report:

EN 1276:2009

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas – Test method and requirements (phase 2, step 1)

Identification of the test laboratory:

Abbott Analytical Ltd
Unit 2, Hickmans Road, Birkenhead, CH41 1JH, United Kingdom

Identification of the client:

Safe Solutions (Safe4) Ltd
Wharton Green, Bostock Road, Winsford, CW7 3BD, United Kingdom

Identification of the sample:

19F/006

Name of the product:

PDSA Shampoo

Batch number/reference and
expiry date (if available):

3671

Date of delivery:

7 June 2016

Storage conditions:

Room temperature in darkness

Product diluent recommended by
the manufacturer for use:

Not disclosed

Active substance(s) and their
concentrations (s) (optional):

Not disclosed

Appearance of the product:

Thick pink liquid

Notes:

- 1) The test results in this report relate only to the sample(s) tested.
- 2) This test report may not be reproduced except in full, adapted, altered or used to create a derivative work, without written approval from Abbott Analytical Ltd.

Test method and its validation:

Method:	Dilution-neutralisation
Neutraliser:	30.0 g/l Polysorbate 80 + 3.0 g/l Lecithin + 1.0 g/l L-histidine + 1.0 g/l L-cysteine (Neutraliser A) or 100.0 g/l Polysorbate 80 + 30.0 g/l Lecithin + 30.0 g/l Tryptone Soya Broth + 5.0 g/l Sodium thiosulphate + 1.0 g/l L-histidine (Neutraliser B)
Neutraliser validation:	Validated in accordance with EN 1276:2009 (5.5.2)

Experimental conditions:

Period of analysis:	14 June 2019 to 18 July 2019
Product test concentration(s):	Neat
Diluent used for product test solution(s):	N/A
Contact time(s):	5 min ± 10 s
Test temperature(s):	20°C ± 1°C
Interfering substance:	0.3 g/l bovine albumin (clean conditions)
Temperature of incubation:	36°C ± 1°C
Identification of the bacterial strain(s) used:	<i>Pseudomonas aeruginosa</i> (NCIMB 10421) <i>Escherichia coli</i> (NCTC 10418) <i>Staphylococcus aureus</i> (NCTC 10788) <i>Enterococcus hirae</i> (NCIMB 8192)

Deviations:

- 1) Additional serial dilutions of *Na* – down to 10⁻² – undertaken at client's request in order to show a greater than 3 lg reduction against each test organism.

Remarks:

- 1) All test conditions are as requested by the client, irrespective of whether these are in accordance with EN 1276:2009 (5.4.2) or EN 1276:2009 (5.5.1.1).
- 2) Products can only be tested at a concentration of 80% or less as some dilution is always produced by adding the test organisms and interfering substance.

Requirements:

The client has requested that the product demonstrates at least a 3 decimal log (lg) reduction against every test organism.

Conclusion:

PDSA Shampoo achieves a greater than 3 lg reduction when tested neat with a contact time of 5 minutes at 20°C under clean conditions against all of the referenced strains of *Pseudomonas aeruginosa*, *Escherichia coli*, *Staphylococcus aureus* and *Enterococcus hirae*.

Report prepared by:

Signed:



Name:

Tony Watson

Position:

General Manager

Date:

19 July 2019

Approved by:

Signed:



Name:

Gareth Bayliss

Position:

Laboratory Manager

Date:

19 July 2019

Results: EN 1276:2009

Test organism:	<i>Pseudomonas aeruginosa</i>	(NCIMB 10421)
Date of test:	14 June 2019	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	B	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	56	$\bar{x} =$	Vc1	61	$\bar{x} =$	Vc1	59	$\bar{x} =$	Vc1	55	$\bar{x} =$
Vc2	63	59.5	Vc2	60	60.5	Vc2	67	63	Vc2	57	56
30 ≤ \bar{x} of N_{V_0} ≤ 160 ?			\bar{x} of A ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of B ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of C ≥ 0.5 × \bar{x} of N_{V_0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} $w_m = 4.45 \times 10^8$; $\lg N = 8.65$ $N_0 = N / 10$; $\lg N_0 = 7.65$ 7.17 ≤ $\lg N_0$ ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no
10^{-6}	>330	>330	
10^{-7}	46	43	

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	N_a ($\bar{x} \times 10$ or $\bar{x} w_m \times 10$)	$\lg N_a$	$\lg R$ ($\lg N_0 - \lg N_a$)
<i>Neat</i>	5 min	10^0	>330	>330	3.95×10^4	4.60	3.05
		10^{-1}	>330	>330			
		10^{-2}	39	40			

Results: EN 1276:2009

Test organism: *Escherichia coli* (NCTC 10418)
 Date of test: 2 July 2019
 Test temperature: 20°C ± 1°C Incubation temperature: 36°C ± 1°C
 Dilution-neutralisation method: Pour plate Number of plates: 1 / ml
 Neutraliser: A Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{v0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	94	$\bar{x} =$	Vc1	91	$\bar{x} =$	Vc1	94	$\bar{x} =$	Vc1	91	$\bar{x} =$
Vc2	89	91.5	Vc2	99	95	Vc2	93	93.5	Vc2	92	91.5
30 ≤ \bar{x} of N_{v0} ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of A ≥ 0.5 × \bar{x} of N_{v0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of B ≥ 0.5 × \bar{x} of N_{v0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of C ≥ 0.5 × \bar{x} of N_{v0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} $w_m = 3.60 \times 10^8$; $\lg N = 8.56$ $N_0 = N / 10$; $\lg N_0 = 7.56$ 7.17 ≤ $\lg N_0$ ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no
10 ⁻⁶	>330	>330	
10 ⁻⁷	34	38	

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	N_a ($\bar{x} \times 10$ or \bar{x} $w_m \times 10$)	$\lg N_a$	$\lg R$ ($\lg N_0 - \lg N_a$)
<i>Neat</i>	5 min	10 ⁰	>330	>330	2.46 × 10 ⁴	4.39	3.17
		10 ⁻¹	254	243			
		10 ⁻²	24	20			

Results: EN 1276:2009

Test organism:	<i>Staphylococcus aureus</i>	(NCTC 10788)
Date of test:	16 July 2019	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	A	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{v0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	59	$\bar{x} =$	Vc1	61	$\bar{x} =$	Vc1	60	$\bar{x} =$	Vc1	56	$\bar{x} =$
Vc2	63	61	Vc2	65	63	Vc2	54	57	Vc2	55	55.5
30 ≤ \bar{x} of N_{v0} ≤ 160 ?			\bar{x} of A ≥ 0.5 x \bar{x} of N_{v0} ?			\bar{x} of B ≥ 0.5 x \bar{x} of N_{v0} ?			\bar{x} of C ≥ 0.5 x \bar{x} of N_{v0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} $w_m = 4.50 \times 10^8$; $\lg N = 8.65$ $N_0 = N / 10$; $\lg N_0 = 7.65$ 7.17 ≤ $\lg N_0$ ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no
10^{-6}	>330	>330	
10^{-7}	44	46	

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	Na ($\bar{x} \times 10$ or $\bar{x} w_m \times 10$)	$\lg Na$	$\lg R$ ($\lg N_0 - \lg Na$)
<i>Neat</i>	5 min	10^0	>330	>330	4.00 × 10 ³	3.60	4.05
		10^{-1}	41	39			
		10^{-2}	4	5			

Results: EN 1276:2009

Test organism:	<i>Enterococcus hirae</i>	(NCIMB 8192)
Date of test:	2 July 2019	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	A	Test conditions: Clean conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: <i>Neat</i>		
Vc1	117	$\bar{x} =$	Vc1	119	$\bar{x} =$	Vc1	124	$\bar{x} =$	Vc1	123	$\bar{x} =$
Vc2	118	117.5	Vc2	127	123	Vc2	122	123	Vc2	130	126.5
30 ≤ \bar{x} of N_{V_0} ≤ 160 ?			\bar{x} of A ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of B ≥ 0.5 × \bar{x} of N_{V_0} ?			\bar{x} of C ≥ 0.5 × \bar{x} of N_{V_0} ?		
<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} $w_m = 4.60 \times 10^8$; $\lg N = 8.66$ $N_0 = N / 10$; $\lg N_0 = 7.66$ 7.17 ≤ $\lg N_0$ ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no
10^{-6}	>330	>330	
10^{-7}	47	45	

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	Na ($\bar{x} \times 10$ or $\bar{x} w_m \times 10$)	$\lg Na$	$\lg R$ ($\lg N_0 - \lg Na$)
<i>Neat</i>	5 min	10^0	>330	>330	3.85×10^3	3.59	4.07
		10^{-1}	40	37			
		10^{-2}	4	3			

Explanations:

V_c	count per ml (one plate or more)
\bar{x}	average of V_{c1} and V_{c2} (1 + 2 duplicate)
\bar{x}_{wm}	weighted mean of \bar{x}
N	number of cells per ml in the test suspension
N_o	number of cells in the test mixture at the beginning of the contact time ($N_o = N / 10$)
N_a	number of survivors per ml in the test mixture at the end of the contact time (before neutralisation or filtration)
R	reduction ($\lg R = \lg N_o - \lg N_a$)
N_v	number of cells per ml in the validation suspension
N_{v_o}	number of cells in the validation mixtures at the beginning of the contact time ($N_{v_o} = N_v / 10$)
A	number of survivors per ml in the experimental conditions control mixture at the end of the contact time
B	number of survivors per ml in the neutraliser or filtration control mixture after 5 minutes
C	number of survivors per ml in the method validation mixture after 30 minutes

All test results have an associated uncertainty of measurement, details of which are available on request.